



URGENT MEDICAL DEVICE CORRECTION

GE Healthcare
3000 N. Grandview Blvd. - W440
Waukesha, WI 53188 USA

June 1, 2022

GEHC Ref. # 34129

To: Chief of Anesthesia
Director of Biomedical / Clinical Engineering
Health Care Administrator / Risk Manager

RE: **Aisys CS², Avance CS², and Avance CS² Pro Anesthesia Systems - Potential reversal of O₂ and Air cylinder pressure transducer connections**

This document contains important information for your product. Please ensure all potential Users in your facility are made aware of this safety notification and the recommended actions. Please retain this document for your records.

Safety Issue

Certain Aisys CS², Avance CS², and Avance CS² Pro Anesthesia Systems manufactured from January 1, 2021 to April 23, 2022 can potentially have the O₂ & Air cylinder pressure transducers reversed. The anesthesia system can display the O₂ cylinder pressure reading as the Air cylinder pressure and the Air cylinder pressure reading as the O₂ cylinder pressure. The correct gases are connected to each supply line, but the screen will display the cylinder pressures as reversed.

In rare instances prolonged hypoxia could result if the oxygen cylinder became depleted when both air and oxygen pipeline gases are unavailable and both cylinders are being utilized but not measured appropriately.

There have been no injuries reported as a result of this issue.

Actions to be taken by Customer /User

You can continue to use your anesthesia system by following the below instructions prior to use:

1. Ensure the System Status menu is visible on the screen by pressing the **System Setup** key and then select **System Status**.
2. Ensure all cylinders are fully closed and the System Status menu displays zero for each cylinder.
3. Open each cylinder valve **one at a time** and ensure the corresponding value displayed on the screen indicates the appropriate cylinder pressure.
4. If the display shows the cylinder pressure for the corresponding open cylinder, you can continue to use your anesthesia system. Your system has passed and is not affected by this issue.
5. If the display shows the cylinder pressure does not correspond to the open cylinder, and if you choose to continue to use your anesthesia system, ensure the following:
 - a. Your facility has reliable gas pipeline supplies and all gas cylinders are full prior to using the machine.
 - b. A backup mode of gas delivery, such as an AMBU bag is available.
 - c. All users of the affected device are alerted of the situation.
 - d. If cylinders are in use, note that the pressure readings for O₂ and Air are reversed.
6. Complete the attached Medical Device Notification Acknowledgement Response form and send to FMI34129.AisysTransducer@ge.com

**Affected
Product
Details**

Aisys CS² Anesthesia Systems: P/N: 1011-9050-000 – GTIN: 00840682102292
Avance CS² and Avance CS² Pro Anesthesia Systems:
P/N: 1009-9050-000 - GTIN: 00840682102322
See attached Appendix for a list of affected systems.

Intended Use

The GE Datex-Ohmeda Anesthesia Systems are intended to provide general inhalation anesthesia and ventilatory support to a wide range of patients (neonatal, pediatric, adult). The device is intended for volume or pressure control ventilation by a clinician qualified in the administration of general anesthesia.

**Product
Correction**

GE Healthcare will correct all affected products at no cost to you. A GE Healthcare representative will contact you to arrange for the correction.

**Contact
Information**

If you have any questions or concerns regarding this notification, please contact GE Healthcare Service at 1-800-437-1171 or your local Service Representative.

Please be assured that maintaining a high level of safety and quality is our highest priority. If you have any questions, please contact us immediately per the contact information above.

Sincerely,



Laila Gurney
Chief Quality & Regulatory Officer
GE Healthcare



Jeff Hersh, PhD MD
Chief Medical Officer
GE Healthcare

Appendix A

Affected Serial Numbers

Avance CS2 and Avance CS2 Pro:

Serials starting with APKA

00101 through 00105, 00107 through 00165, 00167 through 00176, 00178 through 00218, 00220 through 00767, 00776 through 00836, 00838 through 01081, 01084 through 01150, 01152 through 01467, 01469 through 01532, 01537 through 01541, 01543 through 01809, 01811 through 01894, 01921 through 01961, 01965 through 01985, 02009 through 02013, 02016 through 02029

Serials starting with APKB

00101 through 00108, 00110 through 00121, 00123 through 00265, 00269 through 00293, 00295 through 00417, 00422 through 00438, 00441 through 00452, 00462, 00464 through 00473, 00478, 00480 through 00487, 00490 through 00492, 00495 through 00513, 00515 through 00517

APKZ00247, 00506, 02331, 02336, 02337, 02510 through 02569

Aisys CS2:

Serials starting with APWA

00101 through 00176, 00185 through 00697, 00700 through 00763, 00768 through 01013, 01016 through 01216, 01218 through 01223, 01225 through 01596, 01601 through 01683, 01685 through 01805, 01808 through 01960, 01964 through 01999, 02003 through 02021, 02026 through 02168, 02175, 02182 through 02308, 02311 through 02361, 02376 through 02405, 02424 through 02429, 02450 through 02456, 02519 through 02531, 02533 through 02539

Serials starting with APWB

00101 through 00151, 00154, 00155, 00162, 00163, 00169 through 00199, 00201 through 00336, 00338 through 00373, 00378 through 00644, 00647 through 00684, 00686 through 00689, 00691 through 00743, 00746 through 00750, 00754 through 00756, 00770 through 00818, 00823 through 00828, 00838 through 00842, 00905 through 00931, 00944 through 00952

Serials starting with APWZ

03034 through 03037, 03053 through 03056, 03060, 03062 through 03066, 03071 through 03075, 03164 through 03171, 03311 through 03314, 03319, 03320, 03381 through 03415, 03420 through 03422



**MEDICAL DEVICE NOTIFICATION ACKNOWLEDGEMENT
RESPONSE REQUIRED**

Please complete this form and return it to GE Healthcare promptly upon receipt and no later than 30 days from receipt. This will confirm receipt and understanding of the Medical Device Correction Notice.

*Customer/Consignee
Name: _____

Street Address: _____

City/State/ZIP/Country: _____

*Customer Email Address: _____

*Customer Phone Number: _____

We acknowledge receipt and understanding of the Medical Device Correction Notice and have executed the instructions as provided in this notification and below are the results of our testing based on the instructions provided.

Please see the next page to document additional Anesthesia System Serial Number information.

Anesthesia System Serial Number	Pass or Fail	Date of Test
ABCD123456		DD-MMM-YYYY

Please provide the name of the individual with responsibility who completed this form.

Signature: _____

*Printed Name: _____

*Title: _____

*Date (DD/MM/YYYY): _____

*Indicates Mandatory Fields

